



# NUTECH CALIBRATORS & ENGINEERS

(NABL Accredited Calibration Laboratory)

Regd. Office & Lab : 24, GOLD PARK, R. B. CONNECTOR, KOLKATA - 700 107

Mobile : 90515 00089, 90518 20792

e-mail: nutech\_calibrators@rediffmail.com, nutech.calibration700053@gmail.com

info@nutechcalibrators.com, Website : www.nutechcalibrators.com

Format No.: F/NCE/LAB/20		<b>CALIBRATION CERTIFICATE OF :</b>	
Calibration Certificate No.: NCE-AR-2200351-21-22		<b>ABBE REFRACTOMETER</b>	
Date of Calibration: 3-Jan-2022	Issue Date: 5-Jan-2022		
ULR No.: CC330421000000351F	Calibration Due On: 3-Jan-2023		
Location of Calibration: At Lab	Frequency of Calibration: 12 Months		
Issued To: M/s : Dey's Medical Stores Pvt. Ltd. Minerva Garden Complex, D.H Road, Joka, 24 Pgs(S), Kolkata-700104	Instrument Receipt Date: 3-Jan-2022		
Service Request/Order No.:			
Description identification of instrument to be calibrated:			
Name: ABBE Refractometer			
Make: Secor	ID No.: DMJQ-02	Range: --	
Model: ---	Sr. No.: ---	Resolution: -	
Location:		Discipline: Mechanical	
Applicable specification of Item to be calibrated: Accuracy/Permissible Limit: Not Specified.			
Basis Of Calibration:			
S.O.P No.:		Description of Method:	
Standard Followed:		Comparison Method	
Recommended Environmental Condition:			
Temperature: 22°C ± 2°C		Environmental Condition during Calibration:	
Humidity: 50%RH ± 10%RH		Temperature: 20.5 °C	Humidity: 51.7 % RH
<b>CALIBRATION RESULT</b>			
Sl. No.	Solution Use	Result as per I.P	
1	Sorvitol	OK	
2	Glycerine	OK	
Remarks: Statement of conformity not mentioned.			
Note:			
i) This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.			
ii) The certificate issued for industrial purpose only not for commercial activities.			
iii) Partial publication/reproduction of this certificate in any form is not permitted without the written consent of NCE.			
This certificate, issued by Nutech Calibrators & Engineers, refers only for the particular item submitted to calibrate.			

*Ahmed Dewan*

CALIBRATED BY: Ahmed Dewan  
(Calibration Engineer)



APPROVED BY: Amit Banerjee  
C.E.O

*Amit Banerjee*