



NUTECH CALIBRATORS & ENGINEERS

(NABL Accredited Calibration Laboratory)

Regd. Office & Lab : 24, GOLD PARK, R. B. CONNECTOR, KOLKATA - 700 107

Mobile : 90515 00089, 90518 20792

e-mail: nutech_cal2006@rediffmail.com, nutech.calibration700053@gmail.com

info@nutechcalibrators.com, Website : www.nutechcalibrators.com



CC-3304

Format No.: F/NCE/LAB/20		CALIBRATION CERTIFICATE OF : MUFFLE FURNACE			
Calibration Certificate No.: NCE-MF-2200337-21-22		Issue Date: 5-Jan-2022			
Date of Calibration: 3-Jan-2022		Calibration Due On : 3-Jan-2023			
ULR No.: CC330422000000337F		Frequency of Calibration: 12 Months			
Location of Calibration: On Site		Instrument Receipt Date : 3-Jan-2022			
Issued To: M/s : Dey's Medical Stores Pvt. Ltd. Minerva Garden Complex, D.H Road, Joka, 24 Pgs(S), Kolkata-700104					
Service Request/Order No.: -					
Description Identification of instrument to be calibrated:					
Name: Muffle Furnace					
Make: i-therm		ID No.: DMJQ-15		Range: Upto 1200 °C	
Model: -		Sr. No.: -		Resolution: 1 °C	
Location: ICPMS Laboratory		Discipline: Thermal			
Applicable specification of item to be calibrated: Accuracy/Permissible Limit:				Not Specified.	
Basis Of Calibration:					
S.O.P No.: SOP/TC/TH-02			Description of Method:		
Standard Followed:			Comparison Method		
Recommended Environmental Condition:			Environmental Condition during Calibration:		
Temperature:		25°C ± 4°C		Temperature	
Humidity:		50%RH ± 10%RH		Humidity	
		23.6 °C		54.9 % RH	
Traceability: Standards used for calibration are traceable to National / International Standards through ISO/IEC: 17025 accredited laboratory					
Sr. No.	Name of the Standard Instruments Used (Make, Sl. No./Code No.)	Range	Resolution	Calibration Due Date	Traceable to/ Certificate No.
1	Dig Thermometer with Sensor (YD13006356)	[-] 50 to 1372°C	--	23-Aug-22	Transcal (Bangalore) TSC/21-22/6893-1
CALIBRATION RESULT					
Parameter/ Range	Sl.No	Temperature Set at Duc in °C	Average Temp. Observed at Ref. Std. in °C	Error in °C	Measurement Uncertainty (±) in °C
Temperature	1	100	101.12	-1.12	1.60
	2	200	201.52	-1.52	
	3	400	401.89	-1.89	
	4	600	601.96	-1.96	
	5	800	803.22	-3.22	
	6	1000	1004.34	-4.34	
	7	1100	1106.25	-6.25	
Note:					
i) Average of 5 readings had been taken in DUC/STD.					
ii) Measurement Uncertainty reported is at approximately 95% confidence level with k=2, Units of Measurement results & Measurement Uncertainty are same as that of range selected-Unless otherwise indicated.					
iii) NABL-133 guidelines are adopted for use of NABL symbol.					
iv) DUC: Device Under Calibration.					
v) Physical Status Of DUC : Ok					
This certificate, issued by Nutech Calibrators & Engineers, refers only for the particular item submitted to calibrate.					

CALIBRATED BY: 
Ahmed Dewan
(Calibration Engineer)



APPROVED BY: 
Amit Banerjee
(C.E.O)

END OF CERTIFICATE